

# POOL EXACT® EZ TEST SPECIFICATIONS

Menu	Test parameter	Range (ppm)	Best Accuracy % †	Tests
1	Free Chlorine (DPD-1)	0.05 – 12	5	100
2	Total Chlorine (DPD-4)	0.05 – 12	5	100
3	Combined Chlorine (DPD-3) <sup>1</sup>	0.05 – 12	5	100
4	pH	6.0 – 8.5 pH	0.2 pH	50
5	Hardness, Calcium (as CaCO <sub>3</sub> )	19 – 550	10	50
6	Total Alkalinity	11 – 200	7.5	100
7	Cyanuric Acid	3 – 110	9	60
8	Copper (Cu <sup>+2</sup> ) <sup>2</sup>	0.06 – 11.0	2	50
9	Phosphate <sup>2</sup>	0.02 – 5.00	8	50
10	Salt (Chloride) <sup>2</sup>	4 – 400	5	25

<sup>1</sup> Combined Chlorine DPD-3 Test requires Free Chlorine (DPD-1) to be run first.

For use in accordance with U.S. Pat No. 7,333,194, US Pat No 7,491,546 Euro Pat No. 1 725 864 DE FR UK,  
and South African Pat No 2007/0628

† Value provided represents best possible accuracy under laboratory conditions, but may vary throughout the detection range



**waterproof IP67**